scientific value to either of the foreign instruments.

Frank W. Creel,

Director, Statutory Import Programs Staff. [FR Doc. 97–32446 Filed 12–10–97; 8:45 am] BILLING CODE 3510–DS–P

DEPARTMENT OF COMMERCE

International Trade Administration

University of Illinois at Urbana-Champaign; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89–651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th & Constitution Avenue, N.W., Washington, DC.

Docket Number: 97–063. Applicant: University of Illinois at Urbana-Champaign, Urbana, IL 61801. Instrument: (2) Gas Composition Analyzers, Model Epison III. Manufacturer: Thomas Swan & Co., Ltd., United Kingdom. Intended Use: See notice at 62 FR 42237, August 6, 1997.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, was being manufactured in the United States at the time the foreign instrument was ordered (April 2, 1997). Reasons: The foreign instrument provides non-invasive testing of thin film surfaces in feedback control mode.

The National Science Foundation, Center for Interfacial Engineering advised on October 4, 1996 that (1) this capability is pertinent to the applicant's intended purpose and (2) it knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use (comparable case).

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which was being manufactured in the United States at the time of order.

Frank W. Creel.

Director, Statutory Import Programs Staff. [FR Doc. 97–32463 Filed 12–10–97; 8:45 am] BILLING CODE 3510–DS–P

DEPARTMENT OF COMMERCE

International Trade Administration

University of California, Los Alamos; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89–651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 97–055. Applicant: University of California, Los Alamos, NM 87545. Instrument: Single Axis Measuring Machine, Model SIP–550M. Manufacturer: Societe Genevoise d'Instruments de Physique, Switzerland. Intended Use: See notice at 62 FR 41361, August 1, 1997.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides resolution to 0.01 µm with a measuring uncertainty of (0.2+0.4L) µm, where L = length. The National Institute of Standards and Technology advised on November 4, 1997 that (1) this capability is pertinent to the applicant's intended purpose and (2) it knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

Director, Statutory Import Programs Staff.
[FR Doc. 97–32464 Filed 12–10–97; 8:45 am]
BILLING CODE 3510–DS–P

DEPARTMENT OF COMMERCE

International Trade Administration

University of Washington; Notice of Decision on Application for Duty-Free Entry of Scientific Instrument

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89–651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211,

U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

Docket Number: 97–077. Applicant: University of Washington, Seattle, WA 98195–7940. Instrument: Isotope Ratio Mass Spectrometer, Model DELTAplus. Manufacturer: Finnigan MAT, Germany. Intended Use: See notice at 62 FR 52685, October 9, 1997.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides a magnetic sector mass analyzer coupled to a gas chromatograph with in-line combustion of the gas effluent for measurement with samples down to the sub-nanomole level. This capability is pertinent to the applicant's intended purposes and we know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

Director, Statutory Import Programs Staff. [FR Doc. 97–32465 Filed 12–10–97; 8:45 am] BILLING CODE 3510–DS–P

DEPARTMENT OF COMMERCE

International Trade Administration

Export Trade Certificate of Review

AGENCY: International Trade Administration, Commerce ACTION: Notice of revocation of Export Trade Certificate of Review No. 92– 00005.

SUMMARY: The Secretary of Commerce issued an export trade certificate of review to World International Investments Corp. Because this certificate holder has failed to file an annual report as required by law, the Secretary is revoking the certificate. This notice summarizes the notification letter sent to World International Investments Corp.

FOR FURTHER INFORMATION CONTACT:

Morton Schnabel, Acting Director, Office of Export Trading Company Affairs, International Trade Administration, 202/482–5131. This is not a toll-free number.

SUPPLEMENTARY INFORMATION: Title III of the Export Trading Company Act of 1982 ("the Act") (Pub. L. 97–290, 15 U.S.C. 4011–21) authorizes the Secretary of Commerce to issue export trade certificates of review. The regulations implementing Title III ("the